

Adeno Ag

CE

Capilia[™] Adeno Neo



Adeno

Rapid test for detecting Adenovirus antigen

Permitting accurate results; simple and easy to use

Easy to read Black Line

Adopting the latest nanotechnology, "Platinum-Gold Colloid"

Fast test results

Reading time is 3-5 min.

Shared use of extraction buffer in Capilia series

1 sample can be used for 4 types of tests (Flu, Adeno, RSV, hMPV)



more reasonable process,less pain for patients





Collecting specimen

Preparing sample

IVD

Capilia Ade

1 sample can be used for up to 4 test plates!


Please follow package insert with each product about shared use of extraction butter.

eno

To detect Adenovirus antigens in pharyngeal swab, nasal swab, nasal aspirate or keratoconjunctivitis swab.

Test procedure

Method of specimen collection



with a pharyngeal swab.



Nasal swab

a nasal swab.





Collect specimen with a pharyngeal swab.

Keratoconjunctivitis swab


ATTERNAL


Sample application Hold the middle of the tube with the fingers and

dispense 3 drops of the sample.

Sample preparation


liquid.






Soak the swab that collected the specimen in the extraction buffer, and stir well. Then, pinch the tip of the swab firmly with the soft wall of the extraction buffer tube with your fingers and squeeze out the swab. Use this squeezed-out liquid as the sample.

When using a nasal aspirate specimen diluted twofold with physiological saline, add 200 μ L of the specimen to the extraction buffer in the tube, and mix well. Use this mixture as the sample



Firmly attach the nozzle (with a filter) provided in the kit to the top of the extraction buffer tube.



Reference data

Clinical data

The result of the clinical performance evaluation in Japan (Comparison with control products)

| Kind of samples | Sensitivity (%) | | Specificity (%) | Accuracy (%) | Total number |
|-----------------|-----------------|---------|-----------------|----------------|-----------------|
| Pharyngeal swab | 100 | (56/56) | 99.2 (125/126) | 99.5 (181/182) | 182 |

The result of the clinical performance evaluation in Japan (Comparison with RCR)

| Kind of samples | Sensitivity (%) | | Spec | ificity (%) | Accuracy (%) | Total number |
|------------------------------|-----------------|---------|------|-------------|----------------|-----------------|
| Nasal swab | 96.5 | (55/57) | 98.2 | (56/57) | 97.4 (111/114) | 114 |
| Nasal aspirate | 98.1 | (52/53) | 98.3 | (58/59) | 98.2 (110/112) | 112 |
| Keratoconjunctivitis swab | 91.2 | (62/68) | 98.4 | (127/129) | 95.9 (189/197) | 197 |

Product summary

- REF CAAD0370 (20 Tests) Kit contents: Test Plate × 20, Extraction Buffer 0.7 mL × 20, Nozzle×20
- REF CAAD0371 Test Plate (10 Tests) Kit content: Test Plate × 10



Shizuoka, 410-2325 Japan URL : http://www.tauns.co.jp/english/

- Validity: 27 months
- Storage: Store at 2-30°C
- Reading time: 3-5 minutes